



Product Note: B6140681

Calibration Coefficient Conversion From HP-8510 to PNA

The calibration coefficients of Calibration Kits designed for the HP-8510 have to be converted in order to use them on PNA Network Analyzers. Agilent Technologies has a program to convert the coefficients and to generate a file for the PNA Analyzer. The size of the program is too large to e-mail and should be downloaded from the links given below.

The following links can be found by using GOOGLE and entering: **pna cal kit editor** into the search box.

PNA Series Data-Based Cal Kit Editor. Information on the Data-Based **Cal Kit Editor**. This Calibration **Kit Editor**, now part of firmware A.04.06 and above, ...
na.tm.agilent.com/pna.dbcal.html – 19k – Cached – Similar pages

From the **Edit PNA Cal Kits** dialog, click Import **Kit** to load the **Cal Kit** you want ...
Recommended: Also click **Edit PNA Cal Kits** – Save As to save the entire ...
na.tm.agilent.com/pna/help/PNAWebhelp/S3_Cals/ModifyCalKits.htm – 68k – Cached – Similar pages
[More results from na.tm.agilent.com]

VNA **Cal Kit Manager** – edit and download **cal kit** coefficients. ... Support for new **cal kit** features in the **PNA** network analyzers with firmware revision 2.0 ...
www.vnahelp.com/ckman_new2.html -8 Cached - Similar pages

Serving the Microwave Industry since 1985
Solution Partner with Maury Microwave



6501 W. Frye Rd. Chandler, AZ 85226 USA
Tel: (480) 940-0740 Fax: (480) 961-4754
E-mail: sales@icmicrowave.com Website: www.icmicrowave.com

